Utilité du monitoring ambulatoire dans la détection de FA

Bruno Benzaquen • MD FRCPC
Cardiologue

HMR Hôpital Maisonneuve-Rosemont
Centre affilié à l'Université de Montréal

Université
de Montréal

SSVQ
Société des sciences
vasculaires du Québec

Divulgations 2015-2016

- BMS- Conférencier et Consultant
- Bayer- Consultant
- Medtronic- Conferencier

Objectifs

- Décrire l'importante relation entre FA et ACV cryptogénique
- Efficacité du monitoring et survol de quelques nouvelles technologies existantes
- Quant, comment et qui traiter?
- Questions

Le Défi des ACV Cryptogéniques

Types des ACV ischémiques

- Atherothrombotique (25-30%)
ex. MVAS carotides
- Cardioembolique (20%)
ex.. FA, thrombus apex, Végétation
- Lacunaire/Petit Vaisseau (15-20%)
ex.. HTA, Diabète
- Autre (5-10%) ex.. Vasculite

Cryptogénique (25-40%)

ACV Cryptogénique

Pourquoi la FA est si importante

- 30-40% des ACV ischémiques sont classifiés comme cryptogénique, malgré un bilan extensif.
- 50% de ces patients peuvent souffrir d'une FA non-diagnostiquée (Donc 40% de la totalité des ACV ischémiques)
- La détection d'une FA chez ces patients changent entièrement le traitement, d'un antiplaquettaire à une ACO

¹ Adams HP Jr, *Stroke*. Jan 1993; 24; 35-41;

² Camm et al, *European Heart Journal* . 2012; 33, 2719-2747

Pourquoi le monitoring ambulatoire est-il important

- La FA est souvent asymptomatique
- La FA est souvent intermittente et imprévisible
- La durée (fardeau) de la FA est importante
- La FA intermittente et asymptomatique a des conséquences importantes de morbidité et de mortalité

Outils diagnostics pour détecter la FA asymptomatique dans l'ACV cryptogénique

- ECG
- HOLTER
- MONITEURS EXTENES DE LONGUE DUREE
- MONITEUR IMPLANTABLE

ECG

Rendement diagnostic estimé a 2%

Bell C, Kapral M Can J Neurol Sci. 2000 Feb; 27(1):25-31.

HOLTER

HOLTER

Rendement diagnostique estimé a 4%

Sources	Intervention	Duration of monitoring	N	AF Detected (%)
Bansil and Karim (Bansil and Karim, 2004)	CCM	NA	121	4.9
Barthelemy (Barthelemy et al., 2003)	CCM	70 h	52	7.7
Gumbinger (Gumbinger et al., 2012)	CCM	24 h	312	11.9
Kallmunzer (Kallmunzer et al., 2012)	CCM	76 h	271	6.6
Lazzaro (Lazzaro et al., 2012)	CCM	73 h	133	6.0
Rem (Rem et al., 1985)	CCM	48 h	169	2.4
Rizos (Rizos et al., 2010)	CCM	48 h	136	21.3
Rizos (Rizos et al., 2012)	CCM	64 h	496	5.4
Ritter (Ritter et al., 2013)	CCM	72 h	1110	1.3
Sulter (Sulter et al., 2003)	CCM	48 h	27	18.5
Sutarnartpong (Sutarnartpong et al., 2014)	CCM	24 h	204	5.8
Vivanco (Vivanco Hidalgo et al., 2009)	CCM	55 h	465	7.1
Fudji (Fuji et al., 2013)	CCM or Holter	NA	113	11.5
Barthelemy (Barthelemy et al., 2003)	Holter	24 h	55	5.5
Douen (Douen et al., 2008)	Holter	24 h	126	9.5
Gladstone (Gladstone et al., 2014)	Holter	24 h	277	3.2
Gumbinger (Gumbinger et al., 2012)	Holter	24 h	192	1.0
Gunalp (Gunalp et al., 2006)	Holter	24 h	26	42.3
Hornig (Hornig et al., 1996)	Holter	24 h	261	3.8
Jabaudon (Jabaudon et al., 2004)	Holter	24 h	139	5.0
Kessler (Kessler and Kessler, 1995)	Holter	24 h	93	0.0
Lazzaro (Lazzaro et al., 2012)	Holter	24 h	133	0.0
Rizos (Rizos et al., 2012)	Holter	24 h	496	2.8
Ritter (Ritter et al., 2013)	Holter	24 h	1110	0.5
Schaer (Schaer et al., 2004)	Holter	24 h	425	2.1
Shafiqat (Shafiqat et al., 2004)	Holter	24 h	210	2.4
Sobocinski (Doliwa Sobocinski et al., 2012)	Holter	24 h	249	2.0
Stahrenberg (Stahrenberg et al., 2010)	Holter	24 h	224	4.8
Tagawa (Tagawa et al., 2007)	Holter	24 h	299	8.4
Thakkar (Thakkar and Bagarhatta, 2014)	Holter	24 h	52	5.8
Rem (Rem et al., 1985)	Holter	24-48 h	51	3.9
Dangayach (Dangayach et al., 2011)	Holter	48 h	51	23.6
Stahrenberg (Stahrenberg et al., 2010)	Holter	48 h	224	6.4
Schuchert (Schuchert et al., 1999)	Holter	72 h	82	6.1
Stahrenberg (Stahrenberg et al., 2010)	Holter	7 d	224	12.5
Ritter (Ritter et al., 2013)	Holter	7 d	60	1.7

MONITEURS EXTERNES DE LONGUE DUREE

La durée d'enregistrement varie de 7-30 jours

- 2 différentes sortes de technologies:
 - » Enregistrement déclenché par le patient symptomatique
 - » Enregistrement auto-déclenché/enregistrement continu par la machine selon un algorithme précis

MONITEURS EXTERNES DE LONGUE DUREE

- 1) Certains ne peuvent qu'enregistrer des données de 2.5-6 minutes si déclenchés par le patient (symptômes)
- 2) Certains gardent en mémoire tous les battements enregistrés et doivent être analysés par après.
- 3) Certains ont des logiciels qui "auto-detectent" les événements dont la FA
- 4) La capacité de la mémoire, les algorithmes de détection et la qualité des enregistrements varient selon les manufacturiers

Ne pas faire confiance aux symptômes! Faible corrélation entre les symptômes et la présence de FA

- **95% des épisodes de FA étaient asymptomatiques**
- 15% du temps corrélés avec FA
- 45% des patients rapportent des symptômes sans FA
- Autres études démontrent un ratio 12:1 de FA (asymptomatique: symptomatique)

¹ Strickberger A. et al. Relationship between atrial tachyarrhythmias and symptoms. *Heart Rhythm*. 2005;2:125-131.

² Page RL, Wilkinson WE, Clair WK, et al. Asymptomatic arrhythmias in patients with symptomatic paroxysmal atrial fibrillation and paroxysmal supraventricular tachycardia. *Circulation*. 1994;89:224-227.

Alive Cor – Kardia Mobile

Monitoring discontinu-Depend des symptômes

Recorded: Sunday, August 21, 2016 at 1:39:29 AM
Heart Rate: 96 bpm Duration: 30s

AliveCor:

ALIVECOR

Zio Patch- iRhythm Technologies

Monitoring continuu de 14 jours

Cardio-Stat- Incentia

Utilisé au Québec- 7 jours de monitoring continu

INSTITUT UNIVERSITAIRE
DE CARDIOLOGIE
ET DE PNEUMOLOGIE
DE QUÉBEC

Validation of a novel single lead heart rhythm monitor

Isabelle Nault MD, Benoit Plourde MD, Jean-François Sarrazin MD, François Philippon MD, Gilles O'Hara MD, Franck Molin MD, Louis Blier MD, Jean Champagne MD
Institut universitaire de cardiologie et pneumologie de Québec, Québec, Canada

UNIVERSITÉ
LAVALE

Abstract

Background

CardioSTAT® is a new "patch-like" single lead heart rhythm monitor. Recording is made through 2 electrodes positioned in a lead I-like configuration. In a previous study, we validated atrial fibrillation detection compared to 12-lead ECG. In this study, we sought to compare its performance to a standard 24h Holter for arrhythmia detection.

Methods and results

Patients undergoing 24h Holter for control or suspicion of atrial fibrillation (AF) were included in the study. Recording of the standard Holter was performed simultaneously with CardioSTAT recording. The CardioSTAT can monitor up to one week in a continuous manner, however for the purpose of this study, only 24h recordings were performed. Analysis of the data was performed by an electrophysiology technician and interpreted in a blinded fashion by a cardiologist.

One hundred seventy-five monitoring were compared. AF was diagnosed in 60 patients and atrial flutter in 9 patients. Agreement between CardioSTAT and standard Holter was 97.7% for AF detection with kappa=0.95 (95%CI 0.9-0.998) and 95 % for atrial flutter detection with kappa=0.5 (95%CI 0.22-0.78). No significant difference was found in premature atrial and ventricular beat counts, (PAC 463±1414 vs. 536±1896, and PVC 840±2394 vs. 591±2152 for Holter vs CardioSTAT respectively). Mean and minimal heart rate (68±11 vs. 67±11; 47±16 vs. 48±12) were similar between the 2 recordings whereas maximal heart rate was slightly higher in Holter group (116±28 vs 113±27, p=0.004). Longest RR interval (1.87±0.53 vs. 1.85±0.51 seconds, p=0.04). Noise was recorded in 4% of Holter and 8.6% of CardioSTAT.

Conclusion

Preliminary results for validation of a novel heart monitoring device show excellent correlation with the standard Holter for detection of atrial fibrillation. Added value of CardioSTAT—included the possibility of longer monitoring duration (up to 7 days), less cumbersome installation (2 electrodes and no wires) and water resistance (patient can shower with the device).

Disclosure

No conflict of interest related to this study

Background

- Current 24h Holter is widely used for arrhythmia detection in patients with symptoms or for atrial fibrillation detection in asymptomatic patients after a stroke.
- CardioSTAT is a new heart rhythm monitor with continuous recording of a single lead tracing. It is less cumbersome than a regular Holter, it is waterproof, it has a longer recording capacity. Electrodes can be replaced by the patient.
- The aim of this study is to validate the diagnostic accuracy of CardioSTAT compared to a 24h standard Holter.

Methods

- CardioSTAT was installed simultaneously with a 24h Holter in patients undergoing Holters.
- Indication for monitoring was suspicion of atrial fibrillation, assessment of rate / rhythm control in patients known with atrial fibrillation or monitoring after atrial fibrillation ablation.
- Duration of both recordings (CardioSTAT and Holter) was 24h.
- Analysis of tracing was performed by a blinded cardiologist.
- Data was compared using paired-T-test, and kappa correlation.

Results

Kappa correlation for diagnosis of AF: 0.95, SE 0.05, CI 0.9-0.99

	Holter	CardioSTAT	p
# Heart Beats	96411±29790	86257±19484	<0.0001
Min HR	47±16	48±12	ns
Mean HR	68±11	67±11	ns
Max HR	116±28	113±27	0.004
Noise(%)	4.0±4.6	8.6±11.7	<0.0001
PVC (#)	840±2394	591±2394	ns
PAC (#)	463±1414	536±1896	ns
Longest RR	1.87±0.5	1.85±0.5	0.04
AF detection (%)	34%	34%	ns

Concerning noise detection, modification of the device installation was performed after the first 43 patients.

When analysing only patients 44 to 175, noise detection was 5% vs 7% Holter vs CardioSTAT, p=ns

Conclusions

CardioSTAT® shows excellent correlation with 24h Holter for detection of atrial fibrillation, PAC and PVC. Improvement in installation methods reduced noise.

Figure 1 : a- Example of CardioSTAT installation vs b- standard Holter installation

Figure 2: CardioSTAT tracings from a patient with Atrial Fibrillation

Figure 3: CardioSTAT tracings of a patient in sinus rhythm

Spider Flash-Livanova

Monitoring de longue durée- 30 Jours. Autodétection de FA

The NEW ENGLAND JOURNAL of MEDICINE

Atrial Fibrillation in Patients with Cryptogenic Stroke

David J. Gladstone, M.D., Ph.D., Melanie Spring, M.D., Paul Dorian, M.D., Val Panzov, M.D., Kevin E. Thorpe, M.Math., Judith Hall, M.Sc., Haris Vaid, B.Sc., Martin O'Donnell, M.B., Ph.D., Andreas Laupacis, M.D., Robert Côté, M.D., Mukul Sharma, M.D., John A. Blakely, M.D., Ashfaq Shuaib, M.D., Vladimir Hachinski, M.D., D.Sc., Shelagh B. Coutts, M.B., Ch.B., M.D., Demetrios J. Sahlas, M.D., Phil Teal, M.D., Samuel Yip, M.D., J. David Spence, M.D., Brian Buck, M.D., Steve Verreault, M.D., Leanne K. Casaubon, M.D., Andrew Penn, M.D., Daniel Selchen, M.D., Albert Jin, M.D., David Howse, M.D., Manu Mehdiratta, M.D., Karl Boyle, M.B., B.Ch., Richard Aviv, M.B., Ch.B., Moira K. Kapral, M.D., and Muhammad Mamdani, Pharm.D., M.P.H., for the EMBRACE Investigators and Coordinators*

N Engl J Med 2014; 370:2467-2477 | June 26, 2014 | DOI: 10.1056/NEJMoa1311376

Moniteurs externes de longue durée

Rendement diagnostique estimé a 16-20% a 30 jours

Medium and long-term monitoring and AF detection.

Sources	N	AF Definition	Intervention	Duration of monitoring	AF detected (%)
Barthelemy (Barthelemy et al., 2003)	28	30 s	ELR	70 h	14.3
Jabaudon (Jabaudon et al., 2004)	88	"seconds"	ELR	7 d	5.7
Wallman (Wallmann et al., 2007)	127	30 s	ELR	7 d	14.2
Elijovich (Elijovich et al., 2009)	20	30 s	ELR	30 d	20.0
Flint (Flint et al., 2012)	236	30 s	ELR	30 d	7.0
Gaillard (Gaillard et al., 2010)	98	30 s	TTM	30 d	9
Gladstone (Gladstone et al., 2014)	280	30 s	ELR	30 d	16.1
		2.5 min			9.9
Bhatt (Bhatt et al., 2011)	62	30 s	MCOT	28 d	24
		5 min			9
Kamel (Kamel et al., 2013)	20	30 s	MCOT	21 d	0.0
Miller (Miller et al., 2013)	156	30 s	MCOT	30 d	5
Tayal (Tayal et al., 2008)	56	30 s	MCOT	21 d	5

Valeur incrémentale de prolonger la durée de monitoring et la détection de FA

- N = 1491
- ACV aigu/ICT pas de FA connue
- ECG
- HOLTER de 24hr si ECG normal
- Moniteur externe de 7-14 jours si HOLTER normal

1. Jabaudon D. Et al. *Usefulness of Ambulatory 7-Day ECG Monitoring for the Detection of Atrial Fibrillation and Flutter After Acute Stroke and Transient Ischemic Attack* Stroke 2004;35:1647-1651.
2. Cotter, P.E., et al., *Incidence of atrial fibrillation detected by implantable loop recorders in unexplained stroke*. Neurology, 2013 Apr 23;80(17):1546-50
3. Ritter, M.A., et al., *Occult Atrial Fibrillation in Cryptogenic Stroke: Detection by 7-Day Electrocardiogram Versus Implantable Cardiac Monitors*. Stroke, 2013 May;44(5):1449-52.

Moniteur implantable- Reveal Linq (Medtronic)

- Monitoring en continu pour une durée de 2-3 ans
- Programmable
- Compatible avec IRM
- Capabilité d'enregistrement déclenché par patient
- Capabilité d'auto-enregistrement épisodes de FA selon un algorithme préétabli.
- Transmission d'épisodes par Wireless

Reveal linq-Medtronic

Reveal LINQ™ ICM

Wireless

MyCareLink™
Patient Monitor

Cellular

Simplified Insertion
Procedure

Patient
Assistant

Patient Details: Johnson, Ree		Transmission Details		Transfer (Out)	
PID450780321		Last Audit: 26-Jan-2012		Discontinue	
Reveal LINQ™ Date of Implant: 1-Dec-2011					
Overview	Profile	Equipment	History	Schedule	CareAlert Notification
Indication	Date of Birth	Device Information	Patient Phone Number	Follow-Up Physician	
Syncope	01-Jan-1936	Battery Status: OK	955-555-5555 955-777-2340	Smith, Martin, MD	
Patient Reports (Last 30 Reports)			Next Scheduled 30-Day Summary Report: 09-Feb-2012		
Create New Report					
Received	Report Type	Event Summary			
26-Jan-2012 12:14	Full Report	Pause Episode, Symptom Episode			
26-Jan-2012 05:30	Event Report	Pause Episode, Symptom Episode			
7-Jan-2012 08:15	Summary File				
9-Dec-2011 11:53	Full Report

Mobile
Alerts

Streamlined Reports

Reveal-Linq

Minimally Invasive Outpatient Insertion Procedure - No Leads or Fluoroscopy

AF Monitoring with Reveal LINQ: Rhythm (ir-)regularity

Sinus rhythm

Normal variations

AF

Irregular irregularity

Reveal-Linq

Reveal Linq- Fardeau de FA

Battery Status

Good

Episodes (14) >>

Symptom	1
FVT	2
VT	1
Asystole	0
Brady	0
AT	0
AF	10

% of Time AT/AF 27.7 %

Observations (3) >>

- FVT/VT episodes have reached the programmed values, most recent episode #44.
- AT/AF daily burden exceeded programmed threshold.
- V. Rate during AT/AF exceeded programmed threshold in a 24 hour period.

Parameter Settings >>

	Detection	ECG	Interval (Rate)	Duration
FVT	On	On	300 ms (200 bpm)	18/24 beats
VT	On	On	360 ms (167 bpm)	16 beats
Asystole	On	On		3 sec
Brady	On	On	2000 ms (30 bpm)	4 beats
AT/AF	AF Only	On

Quick Look

Episodes

Params

< Reports

Patient

< Session

Print...

Open File...

End Session...

Landmark CRYSTAL AF Study:

CRYptogenic STroke And underLYing Atrial Fibrillation

Objectives:

- Assess ICM vs. Standard Care for detection of AF in cryptogenic stroke patients
 - 6 month endpoint (primary)
 - 12 month endpoint (secondary)
- Determine proportion of patients with underlying AF
- Record actions taken after AF diagnosis

CRYSTAL AF Study Overview

- Patient had **Cryptogenic stroke/TIA**
- Randomized to **SoC or ICM**
- AF defined as ≥ 30 seconds

- **Primary endpoint**
 - AF detection at 6 months
- **Secondary endpoints**
 - AF detection at 12 months
 - AF duration
 - Symptom correlation
 - Physician actions

CRYSTAL AF Study

Patient Inclusion Criteria:

- ≥ 40 years of age
- Cryptogenic stroke (or clinical TIA), with infarct seen on MRI or CT, within the previous 90 days; and no mechanism (including AF) determined after:
 - **12-lead ECG**
 - **24-hour ECG monitoring (e.g. Holter)**
 - **Transesophageal echocardiography (TEE)**
 - **CTA or MRA of head and neck to rule out arterial source**
 - **Screening for hypercoagulable states in patients < 55 years old**

Patient Exclusion Criteria:

- History of AF or Atrial Flutter
- Permanent indication or contraindication for anticoagulation
- Indication for pacemaker or ICD

CRYSTAL AF: At 36 months, 30% of patients in the ICM arm found to have AF¹

8.8x more than standard follow-up arm

AF Detection at 36 Months

CRYSTAL AF Study Results

Reveal ICM superior to SoC for finding AF in patients with a cryptogenic stroke¹

- Over **7x more** patients with AF at 12-months
- At 36 months, **30%** of patients in the Reveal ICM arm found to have AF
- Short-term monitoring not sufficient: median time to AF detection over 12 months was **84 days**
- At 12 months, **97%** of patients with AF prescribed OAC

ESC-Directives de FA (2016)

Recommendations for screening for atrial fibrillation

Recommendations	Class ^a	Level ^b	Ref ^c
Opportunistic screening for AF is recommended by pulse taking or ECG rhythm strip in patients >65 years of age.	I	B	130, 134, 155
In patients with TIA or ischaemic stroke, screening for AF is recommended by short-term ECG recording followed by continuous ECG monitoring for at least 72 hours.	I	B	27, 127
It is recommended to interrogate pacemakers and ICDs on a regular basis for atrial high rate episodes (AHRE). Patients with AHRE should undergo further ECG monitoring to document AF before initiating AF therapy.	I	B	141, 156
In stroke patients, additional ECG monitoring by long-term non-invasive ECG monitors or implanted loop recorders should be considered to document silent atrial fibrillation.	IIa	B	18, 128
Systematic ECG screening may be considered to detect AF in patients aged >75 years, or those at high stroke risk.	IIb	B	130, 135, 157

Qui et quand anticoaguler?

- Tout patient sans contre-indication qui a souffert un ACV/ICT cryptogénique?
- Les études sur le monitoring ont considéré la « FA » significative ayant des épisode de 30 sec ou plus.
- Quel est le fardeau de FA nécessaire pour vraiment augmenter les risques?
- Deux façons de voir les choses

Tirade de Virchow et Directives cliniques

VIRCHOW'S TRIAD

and THROMBUS FORMATION
IN ATRIAL FIBRILLATION

More than 150 years ago, Virchow postulated a triad of factors contributing to **thrombogenesis**. Today, these are active drivers of stroke risk in atrial fibrillation.

The "CCS Algorithm" for OAC Therapy in AF

Consider and modify (if possible) all factors influencing risk of bleeding during OAC treatment (hypertension, antiplatelet drugs, NSAIDs, excessive alcohol, labile INRs) and specifically bleeding risks for NOACs (low eGFR, age ≥ 75, low body weight).¹

Le Fardeau de FA importe-t-il? (ASSERT, nejm 2012)

The NEW ENGLAND JOURNAL of MEDICINE

ORIGINAL ARTICLE

METHODS

We enrolled 2580 patients, 65 years of age or older, with hypertension and no history of atrial fibrillation, in whom a pacemaker or defibrillator had recently been implanted. We monitored the patients for 3 months to detect subclinical atrial tachyarrhythmias (episodes of atrial rate >190 beats per minute for more than 6 minutes) and followed them for a mean of 2.5 years for the primary outcome of ischemic stroke or systemic embolism. Patients with pacemakers were randomly assigned to receive or not to receive continuous atrial overdrive pacing.

mias was 13%. Subclinical atrial tachyarrhythmias remained predictive of the primary outcome after adjustment for predictors of stroke (hazard ratio, 2.50; 95% CI, 1.28 to 4.89; $P=0.008$). Continuous atrial overdrive pacing did not prevent atrial fibrillation.

ESC-Directives de FA (2016)

Stroke Algorithm

Questions

- Dans l'ère des NACO, chez des patients bien choisis le risque de saignement est faible et les bénéfices supérieurs, devons nous anticoaguler tout les ACV cryptogéniques?
- Devons nous chercher la FA chez des patients avec des score CHADS2 élevés même sans antécédents d'ACV?
- Corrélation entre le fardeau FA/score CHADS pour mieux guider les traitement optimal.

Conclusions

- La FA est responsable probablement de 20% des ACV ischémiques et de 40-50% des ACV cryptogéniques donc 40-50% de la totalité des ACV ischémiques avec des répercussions très importantes.
- Peu de rendement diagnostique des ECG, HOLTER, meilleur des Moniteurs a long terme (selon la technologie utilisée)
- Rendement supérieur des moniteurs implantables qui devraient être le gold-standard pour le diagnostique de FA.
- Il faut encore utiliser le jugement clinique pour individualiser le traitement et décider quand, qui et comment on anti coagule.

Questions?